

Changing Batteries in the Mobile Power Unit (MPU)



Every 6 months the batteries in the MPU need to be changed.

1. Gather 3 new “AA” batteries and a coin.
2. On the bottom of the MPU, use a coin to open the battery compartment.



3. Change the batteries. Ensure the red ribbon is located below the batteries.



4. Replace the cover.

If you have any questions, please call the VAD Office at 612-863-5638 and ask to speak to a VAD Coordinator.